

12.410 Van Veen Grab – 2500 cm²

Manual



Research Equipment
Limnology • Oceanography • Hydrobiology

	Manual for Van Veen grab 2500 cm ²	Model no. 12.410
	Caution This Van Veen grab is very dangerous in unskilled hands and you must take serious precautions to avoid accidents and injury to personnel. KC Denmark is not, and cannot be held, responsible for any damage(s) made to equipment or to operators who ignore safety precautions or because of misuse or wrong operation.	
	Preparation:	
1	Handle the grab with care to prevent injury to personnel. Fingers and hands may be stuck at the arms or by the jaws.	
2	Inspect the thimbles and talurit ferrules for any damages or weakness. Replace if necessary. Otherwise, there may be a risk of loosing the grab in the sea.	

3	Caution Inspect the wire locks at both ends. It has to lock the wire to avoid the grab from hanging askew.	
4	Place the grab on a plain, stable surface. Force down both arms to horizontal position so the grab is open.	
5	For larger performance, you can add a maximum of 4 lead weights, 12,5 kg each (optional). Fasten the lead weights (optional) beneath the drainage holes. Use a M10 nut placed at the inner side of the grab.	
6	The locking pawl is shown in open (non-secured) position	

Press the locking pawl upwards to lock the arms. 7 From the upper side you can see, if it is locked 8 properly. Hold the steel wire straight to hold the pawl in its lock-position. The grab is ready and attach the wire for deploying to the eye of the wire. The photo shows a smaller grab with thinner wire and accessories for lifting the grab. 9 Caution The wires on the grab must be kept straight to avoid an unintended release of the locking mechanism. Recommended wire for the deployment: Minimum 6-8 mm 10 Align the wire at the inner side (towards the middle of the grab) of the two guiding rails; otherwise, the wire may be stuck at the arms upon release of the grab.

_			
11	Lower the grab to the seabed at a reasonable speed. When the grab hits the seabed, the wire will slacken and the locking pawl is released. Upon retrieving the grab, the shovels will close, and the sample is done.		
	Emptying the grab		
12	The grab has 4 lids and turning the knob counter clockwise, the appropriate lid can be open and you can take out a small sample without emptying the whole sediment sample. Emptying the whole content of the grab is done by forcing the arms of the grab into horizontal position. Caution When handling the grab, take care to avoid any injury. See item #1		
	Maintenance		
	i idiileeiidi		
13 A	All parts of the Van Veen grab can be rinsed using salt water or fresh water. Regular cleaning with fresh water is recommended and all moveable parts must be moved individually to ensure all dirt has been removed. A. The locking pawl for the releaser and the hinges.		

13 C C. The wheels at the end of the arms.



Technical	Technical information		
Materials:			
All metal parts:	AISI 316 stainless steel. Finish: Electro polish.		
Wheels for wire:	Brass		
Wire:	Stainless steel wire		
Weights:	Lead, 12,5 kg each. Maximum mounting: 4 pcs.		
Dimensions and weight:			
Overall dimensions, open. L x W x H:			
Overall dimensions, closed. L x W x H:			
Weight, exclusive lead weights:	125 kg net		
Weight, inclusive 4 lead weights:	175 kg net		
Sieves for the drainage:	Stainless steel, mesh size 500 µ		
Operational depth:			
Full oceanographic depth:	>6000 m (No limits).		

Rev.: October 7, 2021 - lkj



Research Equipment
Limnology • Oceanography • Hydrobiology